



Baltimore, MD
410-354-3300

Union City, CA
510-489-6300

Santa Clara, CA
408-748-3585

Austin, TX
512-287-2500

www.metlabs.com

MET Laboratories, Inc.

11/12/2018

William Holt
Dakota Ultrasonics
1500 Green Hills Rd., #107
Scotts Valley, CA 95066

Dear William Holt:

It is our pleasure to inform Dakota Ultrasonics that the CMX USB as defined in MET Laboratories' Project number 101405 has completed the following tests performed at MET Laboratories, Inc.

Dept	Title	Status	Date
EMC	EN 61326-2-6	Completed	10/29/2018
EMC	Emissions (Class A)	Completed	10/31/2018
EMC	Radiated Emissions	Compliant	10/31/2018
EMC	Immunity (Table 1)	Completed	10/29/2018
EMC	Electrostatic Discharge	Compliant	10/31/2018
EMC	Radiated Immunity	Compliant	11/2/2018
EMC	FCC Part 15 Subpart B (Class A)	Completed	10/29/2018
EMC	Radiated Emissions - 15.109	Compliant	10/31/2018
EMC	ICES-003 (Class A)	Completed	10/29/2018
EMC	Radiated Emissions	Compliant	10/31/2018

For specific details regarding the extent of the testing and the configuration of the product tested, please consult the MET Laboratories' Detailed Test Report.

This document is for informational purposes only. Please contact us with any questions or comments.

Document Approved by:

John Mason
Director, EMC Business Line

Steve Pitta
Director, ENV Business Line