



MET Laboratories, Inc.

Baltimore, MD
410-354-3300

Union City, CA
510-489-6300

Santa Clara, CA
408-748-3585

Austin, TX
512-287-2500

www.metlabs.com

January 10, 2018

Mr. William Holt
Dakota Ultrasonics
1500 Green Hills Road #107
Scotts Valley, CA 95096

Dear Mr. William Holt,

Congratulations! It is our pleasure to inform Dakota Ultrasonics that the DFX-7 USB, as defined in the MET Laboratories' Project No. 96296, has successfully completed and is compliant with the following tests performed at MET Laboratories, Inc.:

- ***Title 47 of the CFR, Part 15, Subpart B for a Digital Device, Class B***
- ***EN 61326-1: 2013*** – Emissions:
 - Radiated Emissions, Class B
- ***EN 61326-1: 2013*** – Immunity:
 - EN 61000-4-2: 2009 – ESD
 - EN 61000-4-3: 2006 +A1:2008 +A2:2010 – Radiated Immunity

For specific details regarding the extent of the testing and the configuration of the product tested, please consult the MET Laboratories' Detailed Test Report.

Please contact me with any questions or comments.

Sincerely,

John Mason
Director, EMC Lab