



MET Laboratories, Inc.

Baltimore, MD
410-354-3300

Union City, CA
510-489-6300

Santa Clara, CA
408-748-3585

Austin, TX
512-287-2500

www.metlabs.com

May 25, 2016

William Holt
Dakota Ultrasonics
1500 Green Hills Road #107
Scotts Valley, CA 95066

Dear Mr. Holt,

Congratulations! It is our pleasure to inform Dakota Ultrasonics that the DFX-8 and Max II, as defined in the MET Laboratories' Project No. 88356, has successfully completed and is compliant with the following tests performed at MET Laboratories, Inc.:

- **Title 47 of the CFR, Part 15, Subpart B for a Digital Device, Class A**
- **EN 61326-1: 2013** – Emissions:
 - Radiated Emissions, Class A
 - Conducted Emissions Voltage, Class A
- **EN 61326-1: 2013** – Immunity:
 - EN 61000-4-2: 2009 – ESD
 - EN 61000-4-3: 2006 +A1:2008 +A2:2010 – Radiated Immunity
 - EN 61000-4-4: 2012 – EFT/B Immunity
 - EN 61000-4-5: 2006 – Surge
 - EN 61000-4-6: 2009 – Conducted Immunity
 - EN 61000-4-11: 2004 – Voltage Interruption / Dips

For specific details regarding the extent of the testing and the configuration of the product tested, please consult the MET Laboratories' Detailed Test Report.

Please contact me with any questions or comments.

Sincerely,

Asad Bajwa
Director, EMC Lab